

Electrical apparatus for the detection and measurement of combustible gases — Performance requirements for Group I apparatus indicating up to 100 % (V/V) methane in air

The European Standard EN 50056:1998 has the status of a
British Standard

ICS 13.320

National foreword

This British Standard is the English language version of EN 50056:1998. It supersedes BS EN 50056:1991.

The 1998 edition of the European Standard incorporates some minor modifications to bring it into alignment with the New Approach Directive on Potentially Explosive Atmospheres (ATEX). There are no changes to the technical provisions.

The UK participation in its preparation was entrusted by Technical Committee GEL/31, Electrical apparatus for explosive atmospheres, to Subcommittee GEL/31/19, Gas detectors, which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible European committee any enquiries on the interpretation, or proposals for change, and keep the UK interests informed;
- monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this subcommittee can be obtained on request to its secretary.

Cross-references

The British Standards which implement international or European publications referred to in this document may be found in the BSI Standards Catalogue under the section entitled “International Standards Correspondence Index”, or by using the “Find” facility of the BSI Standards Electronic Catalogue.

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, the EN title page, pages 2 to 5 and a back cover.

This British Standard, having been prepared under the direction of the Electrotechnical Sector Committee, was published under the authority of the Standards Committee and comes into effect on 15 January 1999

© BSI 01-1999

Amendments issued since publication

Amd. No.	Date	Text affected

English version

Electrical apparatus for the detection and measurement of combustible gases — Performance requirements for Group I apparatus indicating up to 100 % (V/V) methane in air

Appareils électriques de détections et de mesure des gaz combustibles — Règles de performances des appareils du Groupe I pouvant indiquer jusqu'à 100 % (V/V) de méthane dans l'air

Elektrische Geräte für das Aufspüren und die Messung brennbarer Gase — Anforderungen an das Betriebsverhalten von Geräten der Gruppe I mit einem Meßbereich bis zu 100 % (V/V) Methan in Luft

This European Standard was approved by CENELEC on 12 September 1996.

CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Electrotechnische Normung

Central Secretariat: rue de Stassart 35, B-1050 Brussels

Foreword

This second edition of the European Standard was prepared by SC 31- 9, Electrical apparatus for the detection and measurement of combustible gases, to be used in industrial and commercial potentially explosive atmospheres, of Technical Committee CENELEC TC 31, Electrical apparatus for explosive atmospheres, on the basis of EN 50056:1991, its amendment A1:1995 and a second amendment. This second amendment was approved by CENELEC on 1996-12-09 for incorporation into a new edition of EN 50056.

This European Standard replaces EN 50056:1991 + A1:1995.
The following dates were fixed:

- latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 1999-01-01
- latest date by which the national standards conflicting with the EN have to be withdrawn (dow) 1999-01-01

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and covers essential requirements of EC Directive 94/9/EC.

Contents

		Page
1	Scope	3
2	Normative references	3
3	Definitions	3
4	General requirements	3
5	Performance requirements	3
5.1	General	3
5.2	Unpowered storage	3
5.3	Calibration curve (not applicable to alarm-only apparatus)	3
5.4	Drift (continuous duty apparatus)	3
5.5	Drift (spot reading apparatus)	3
5.6	Alarm	3
5.7	Temperature	3
5.8	Pressure	3
5.9	Humidity	3
5.10	Air speed	3
5.11	Pumping rate	3
5.12	Orientation	3
5.13	Vibration (applicable only to machine-mounted apparatus)	4
5.14	Drop test (applicable to portable apparatus and remote sensors)	4
5.15	Warm-up time (not applicable to spot-reading apparatus)	4
5.16	Time of response (not applicable to spot-reading apparatus)	4
5.17	Minimum time of operation (spot reading apparatus)	4
5.18	High gas concentration above the measuring range	4
5.19	Battery capacity	4
5.20	Power supply variations	4
5.21	Power supply interruptions, voltage transients and step changes of voltage	4
5.22	Addition of sampling probe	4
5.23	Dust	4
5.24	Poisons and other gases	5
5.25	Electromagnetic immunity	5
6	Field calibration kit	5
7	Information for use	5

1 Scope

1.1 This European Standard specifies performance requirements for Group I (as defined in EN 50054) portable, transportable and fixed apparatus for the detection and measurement of methane concentrations in mine air. The apparatus, or parts thereof, are intended for use in mines susceptible to firedamp and shall meet the general requirements and test methods specified in EN 50054.

1.2 This European Standard is restricted to apparatus intended for the detection and measurement of volume ratios of methane in air from 0 % (V/V) up to 100 % (V/V).

NOTE Apparatus covered by this European Standard will normally be intended to operate in volume ratios greater than 5 % (V/V).

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to the European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

EN 50054:1998, *Electrical apparatus for the detection and measurement of combustible gases — General requirements and test methods*.

3 Definitions

For the purposes of this European Standard, the definitions given in EN 50054 apply.

4 General requirements

The apparatus shall comply with the general requirements specified in clause 4 of EN 50054.

5 Performance requirements

5.1 General

The normal conditions for test are specified in subclause 5.3 of EN 50054. Compliance shall be determined in accordance with the test methods specified in subclause 5.4 of EN 50054.

5.2 Unpowered storage

After being submitted to the conditions specified in subclause 5.4.2 of EN 50054, the apparatus shall meet the requirements specified in 5.3 to 5.25 of this European Standard.

5.3 Calibration curve (not applicable to alarm-only apparatus)

Each of the three indications (after correction using the manufacturer's calibration curve, if necessary) obtained for each of the four gas volume ratios shall not differ from these volume ratios by more than $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater.

5.4 Drift (continuous duty apparatus)

The medium term variation shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater, in air and in the standard test gas.

5.5 Drift (spot reading apparatus)

The variation of the indication shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater, in air and in the standard test gas.

5.6 Alarm

The alarm shall operate during every cycle of the test. If a latching alarm is provided, the manual reset action shall be checked during every cycle.

5.7 Temperature

The variation of the indication from that at 20 °C, over the temperature range:

- a) -10 °C to 0 °C , shall not exceed $\pm 7\%$ (V/V) methane or $\pm 15\%$ of the indication; and
- b) 0 °C to $+40\text{ °C}$, shall not exceed $\pm 5\%$ (V/V) methane or $\pm 10\%$ of the indication.

whichever is the greater.

5.8 Pressure

The variation of the indication from that at 101,3 kPa, over a pressure range of 92 kPa to 115 kPa, shall not exceed $\pm 7,5\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater.

5.9 Humidity

At 40 °C, the variation of the indication from that at 55 % relative humidity (r.h.) over a humidity range of 5 % r.h. to 90 % r.h., shall not exceed $\pm 5\%$ (V/V) methane or $\pm 10\%$ of the indication, whichever is the greater.

5.10 Air speed

The variation of the indication over a range of air speeds from 0 m/s to 6 m/s shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater.

5.11 Pumping rate

The variation of the indication shall not exceed $\pm 3\%$ (V/V) methane or 5 % of the variation, whichever is the greater.

5.12 Orientation

5.12.1 Portable apparatus

The variation in the indication shall not exceed $\pm 5\%$ (V/V) methane or $\pm 10\%$ of the indication, whichever is the greater.

5.12.2 Fixed and transportable apparatus

The variation in the indication shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater.

5.13 Vibration (applicable only to machine-mounted apparatus)

During the vibration test, the apparatus shall not suffer any loss of function and shall not give a false alarm or fault signal. The apparatus shall not suffer damage resulting in hazard or loss of function.

At the conclusion of the vibration test and after the apparatus sensor has then been exposed to clean air followed by the standard test gas, the variation of the indication from that determined prior to the test shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater.

5.14 Drop test (applicable to portable apparatus and remote sensors)

The apparatus shall not suffer damage resulting in hazard or loss of function.

The variation of the indication shall not exceed $+ 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater.

5.15 Warm-up time (not applicable to spot-reading apparatus)

5.15.1 Fixed and transportable apparatus

The apparatus shall warm-up in clean air to indicate zero to within $\pm 3\%$ (V/V) methane, in a time not exceeding 5 min, or longer if specified by the manufacturer. The apparatus shall warm-up in the standard test gas to give a final indication to within $\pm 3\%$ (V/V) methane in a time not exceeding 5 min, or longer if specified by the manufacturer.

5.15.2 Continuous duty portable apparatus

The apparatus shall warm-up in clean air to indicate zero to within $\pm 3\%$ (V/V) methane in a time not exceeding 60 s. The apparatus shall warm-up in the standard test gas to give a final indication to within $\pm 3\%$ (V/V) methane in a time not exceeding 60 s.

5.16 Time of response (not applicable to spot-reading apparatus)

The time of response t_{50} shall not be greater than 10 s, and t_{90} shall not be greater than 30 s.

The specified times of response shall apply to apparatus in the as-supplied condition and without optional accessories attached to the sensors for special purposes, e.g. collecting cones, weather protection.

5.17 Minimum time of operation (spot reading apparatus)

For apparatus without probe or sample line, the indication shall reach 90 % of the final value in a time not exceeding 15 s.

For aspirated apparatus using a sample line or probe length of not more than 3 m, an additional 10 s is permitted.

5.18 High gas concentration above the measuring range

Not applicable.

5.19 Battery capacity

5.19.1 Battery powered portable continuous duty apparatus

The variation shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater, at the end of the 8 h or 10 h period as appropriate.

At the end of the further 15 min following the indication of low battery condition, the variation shall not exceed $\pm 6\%$ (V/V) methane or $\pm 10\%$ of the indication, whichever is the greater.

5.19.2 Battery-powered portable spot-reading apparatus

The variation shall not exceed $\pm 3\%$ (V/V) methane, or $\pm 5\%$ of the indication, whichever is the greater, at the end of 200 operations.

After a further 10 operations following the indication of low battery condition, the variation shall not exceed $\pm 6\%$ (V/V) methane or $\pm 10\%$ of the indication, whichever is the greater.

5.20 Power supply variations

5.20.1 A.C. powered apparatus

The variation of the indication shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater.

5.20.2 External d.c. powered apparatus

The variation of the indication shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater.

5.20.3 Battery powered apparatus

The variation of the indication shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater, when the apparatus is subjected to a change from the maximum terminal voltage of a new or fully charged battery to the minimum recommended operating voltage of that battery as determined by the built-in battery condition indicator.

For spot reading apparatus, it shall be possible to carry out at least 10 measurements at the minimum operating voltage within the above stated tolerances.

5.21 Power supply interruptions, voltage transients and step changes of voltage

The apparatus shall not yield spurious alarms when the specified interruption, voltage transient or step changes of voltage occurs.

5.22 Addition of sampling probe

The addition of a sampling probe shall not introduce a variation of the indication exceeding $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is greater.

5.23 Dust

The functioning of the apparatus shall not be impaired by dust.

For apparatus where the atmosphere is sampled by diffusion, a 50 % reduction of the cross-sectional area of the apparatus inlet shall not produce a variation of indication exceeding $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater.

5.24 Poisons and other gases

5.24.1 Poisons

Not applicable.

5.24.2 Other gases

The variation of the indication shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater.

5.25 Electromagnetic immunity

When subjected to the electromagnetic immunity test, the variation of the indication shall not exceed $\pm 3\%$ (V/V) methane or $\pm 5\%$ of the indication, whichever is the greater. The apparatus shall suffer no loss of function nor spurious alarm.

6 Field calibration kit

The meter or output indication observed during the use of the field calibration kit in a manner corresponding to the manufacturer's instruction, shall not differ from the correct indication by more than $\pm 5\%$ of the measuring range.

7 Information for use

It shall be verified that the contents of the instruction manual are in accordance with the requirements specified in EN 50054.

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: 020 8996 9000. Fax: 020 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: 020 8996 9001. Fax: 020 8996 7001.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: 020 8996 7111. Fax: 020 8996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: 020 8996 7002. Fax: 020 8996 7001.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager. Tel: 020 8996 7070.